

SPECIFIC SCALES ON PDP FOR AIR

Measuring tube must be in Polysulfon

Graduated scales in Nm³/h, absolute pressure, temperature 20 °C

PDP	Code	1 Bar ABS Nm ³ /h	Code	2 Bar ABS Nm ³ /h	Code	3 Bar ABS Nm ³ /h
2.5	731 900	0.2 - 0.95	731 908	0.25 - 1.3	731 916	0.3 - 1.6
5	731 901	0.5 - 1.9	731 909	0.7 - 2.7	731 917	0.8 - 3.4
10	731 902	0.8 - 3	731 910	1 - 4.2	731 918	1.2 - 5.4
15	731 903	1.4 - 5.6	731 911	2 - 8	731 919	2 - 10
20	731 904	1.5 - 7	731 912	2 - 10	731 920	3 - 13
30	731 905	1.5 - 11	731 913	2 - 15	731 921	2.5 - 18
50	731 906	3 - 18	731 914	4 - 25	731 922	5 - 30
100	731 907	6 - 30	731 915	8 - 44	731 923	10 - 54

PDP	Code	4 Bar ABS Nm ³ /h	Code	5 Bar ABS Nm ³ /h	Code	6 Bar ABS Nm ³ /h
2.5	731 924	0.4 - 1.9	731 932	0.4 - 2.1	731 940	0.5 - 2.4
5	731 925	1 - 3.8	731 933	1.2 - 4.2	731 941	1.2 - 4.6
10	731 926	1.4 - 6.4	731 934	1.6 - 7	731 942	1.6 - 7.4
15	731 927	3 - 12	731 935	3 - 13	731 943	3 - 14
20	731 928	3 - 15	731 936	4 - 17	731 944	4 - 18
30	731 929	3 - 22	731 937	3 - 24	731 945	4 - 26
50	731 930	5 - 35	731 938	6 - 40	731 946	6 - 44
100	731 931	12 - 62	731 939	12 - 70	731 947	15 - 75

PDP	Code	7 Bar ABS Nm ³ /h	Code	8 Bar ABS Nm ³ /h	Code	9 Bar ABS Nm ³ /h
2.5	731 948	0.5 - 2.5	731 956	0.5 - 2.7	731 964	0.6 - 2.9
5	731 949	1.2 - 5	731 957	1.4 - 5.4	731 965	1.4 - 5.8
10	731 950	2 - 8	731 958	2 - 8.8	731 966	2 - 9
15	731 951	3.5 - 15	731 959	3.5 - 16.5	731 967	4 - 17
20	731 952	4 - 20	731 960	5 - 21	731 968	5 - 23
30	731 953	4 - 28	731 961	4 - 30	731 969	4 - 33
50	731 954	8 - 48	731 962	8 - 50	731 970	8 - 54
100	731 955	15 - 80	731 963	15 - 85	731 971	20 - 90

PDP	Code	10 Bar ABS Nm ³ /h
2.5	731 972	0.6 - 3.2
5	731 973	1.6 - 6.4
10	731 974	2 - 10
15	731 975	4 - 19
20	731 976	5 - 25
30	731 977	5 - 35
50	731 978	10 - 60
100	731 979	20 - 100

Measuring tube must be in Polysulfon

Any modification applies without notice.

BAMO MESURES

22, Rue de la Voie des Bans - Z.I. de la Gare - 95100 ARGENTEUIL

Tél : (+33) 01 30 25 83 20 - Web : www.bamo.fr

Fax : (+33) 01 34 10 16 05 - E-mail : info@bamo.fr

SPECIFIC SCALES FOR
PDP

28-11-2012

731 I1 04 A

DEB

731-04/1

731

Graduated scales: HCl 30 – 33 %

Measuring tube
may be in
PVC
or Polysulfon

PDP	Code	HCL 30 - 33 % l/h
2.5	731 980	2.5 - 20
5	731 981	5 - 40
10	731 982	10 - 85
15	731 983	15 - 125
20	731 984	20 - 170
30	731 985	30 - 260
50	731 986	50 - 425
100	731 987	100 - 850

Graduated scales: NaOH 30 %

Measuring tube
may be in
Polysulfon

PDP	Code	NaOH 30 % l/h
2.5	731 988	0.2 - 5
5	731 989	1 - 14
10	731 990	3 - 35
15	731 991	3 - 55
20	731 992	5 - 80
30	731 993	6 - 130
50	731 994	10 - 250
100	731 995	40 - 600

Graduated scales: NaOH 50 %

Measuring tube
may be in
Polysulfon

PDP	Code	NaOH 50 % l/h
15	731 996	0.5 - 10
20	731 997	0.5 - 16
30	731 998	1 - 33
50	731 999	2 - 80

Any modification applies without notice.

BAMO MESURES

22, Rue de la Voie des Bans - Z.I. de la Gare - 95100 ARGENTEUIL

Tél : (+33) 01 30 25 83 20 - Web : www.bamo.fr

Fax : (+33) 01 34 10 16 05 - E-mail : info@bamo.fr

**SPECIFIC SCALES FOR
PDP**

28-11-2012

731 I1 04 A

DEB

731-04/2